



**Saybolt**

A CORE LABORATORIES COMPANY

**FAST TO THE POINT.**

## CALIBRATION CERTIFICATE

Report. No: 102/10481-0/08

### ROLLING BALL VISCOSIMETER

Manufacturer: Kittiwake  
Instrument no: 10889  
Ref. Oils: Fuel oils

Ref. viscosimeter	Ref. Viscosity (cSt)	Measured Visc. (cSt)	Deviation (cSt)	Acceptable deviation (cSt)
CF-150-570A	18.60	20.4	1.7	± 2.3
CF-300-66	173.7	181.0	7.3	± 7.8
CF-350-73	383.4	397.1	13.7	± 17.3
CF-400-4178	723.4	737.8	14.4	± 32.6

### DENSITY METER

Manufacturer/Model: Kittiwake / AS-K1-301  
Instrument no.: 10910  
Ref. Oils: Diesel oil (low visc), Fuel Oil (High Visc)  
Ref. Densimeter: DMA 45 S1, No.: 252638

Hydrometer no.	Ref. density (g/ml)	Measured Dens. (g/ml)	Deviation (g/ml)	Acceptable deviation (g/ml)
08/189074	0.8560	0.856	0.000	± 0.001
07/951066	0.8560	0.856	0.000	± 0.001
07/859066	0.9590	0.959	0.000	± 0.001

Saybolt Danmark A/S, March 14, 2008

Jan Achmann



#### SAYBOLT DANMARK A/S

Fyrtårnvej 11, 2300 København S, Danmark  
Tel. +45 3295 3132, Fax +45 3295 3134  
Internet: www.saybolt.dk  
E-mail: saybolt.denmark@corelab.com

All our activities are carried out under our general terms and conditions  
Reg.nr A/S 217496

